California Dept of Transportation

MPQP

AC, PCC, Rubber & Other Batch Operations

Report Number 10-15-2 | calib date 9-14-2015 | district 01 | 109 by G. Thorpe | plant controller MPAQ Automation

plant type PCC Batch | plant owner NOR-CAL | plant location Shell Ln. WIllits

scale acceptable to 8000 | useable scale capacity 10000 | largest noted error, for this scale 0

scale number 1 | material type cement | graduation size 5.0

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
1	2000	0	2000	always zero for 1st step	2000	2000	0.0	0.0	Acceptable	
2	4000	0	4000	0	4000	4000	0.0	0.0	Acceptable	
3	6000	0	6000	0	6000	6000	0.0	0.0	Acceptable	
4	8000	0	8000	0	8000	8000	0.0	0.0	Acceptable	

scale acceptable to 42000 | useable scale capacity 50000 | largest noted error, for this scale 1 Grad

scale number 2 | material type aggregate | graduation size 20

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
1	12000	0	12000	always zero for 1st step	12000	12000	0	0	Acceptable	
2	12000	12000	24000	0	24000	24000	0	0	Acceptable	
3	12000	21000	33000	0	33000	32980	-20	-1	Acceptable	
4	12000	29980	41980	0	41980	41980	0	0	Acceptable	

scale acceptable to 5000 | useable scale capacity 5000 | largest noted error, for this scale

scale number 3 | material type water | graduation size 62.55

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
1	4162		4162	always zero for 1st step	4162	4124	-38	-.6075	Acceptable	491 Gal Metered 494.5 Delivered O.K.
2	4162		4162		4162	4132	-30	-.4796	Acceptable	491 Gal Metered 495.4 Delivered O.K.

scale acceptable to | useable scale capacity | largest noted error, for this scale

scale number | material type | graduation size

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
				always zero for 1st step						

Test Notes

Plant Certified for State Use.
Water Meter O.K. Span 0905
Interlocks O.K.
Printer O.K.
Barteliomie Scale