California Test **109**

AC, PCC, Rubber & Other Batch Operations

Report Number 11-16 | calib date 04/26/2011 | district 01 | 109 by L. LEE | plant controller SELTEC 920i / RICE LAKE

plant type A/C BATCH PLANT "A" | plant owner MERCER FRASER | plant location SANDY PRARIE ROAD FORTUNA

scale acceptable to 10000 | useable scale capacity 6000 | largest noted error, for this scale 1 GRAD

scale number 1 | material type aggregate | graduation size 10

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
1	1000	0	1000	always zero for 1st step	1000	1000	0.0	0.0	Acceptable	
2	2000	0	2000	0	2000	2000	0.0	0.0	Acceptable	
3	3000	0	3000	0	3000	3000	0.0	0.0	Acceptable	
4	4000	0	4000	0	4000	4000	0.0	0.0	Acceptable	
5	5000	0	5000	0	5000	5010	10.0	1.0	Acceptable	
6	6000	0	6000	0	6000	6010	10.0	1.0	Acceptable	

scale acceptable to 500 | useable scale capacity 400 | largest noted error, for this scale 1 GRAD

scale number 2 | material type asphalt | graduation size 1

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
1	100	0	100	always zero for 1st step	100	100	0	0	Acceptable	
2	200	0	200	0	200	200	0	0	Acceptable	
3	300	0	300	0	300	301	1	1	Acceptable	
4	400	0	400	0	400	401	1	1	Acceptable	

scale acceptable to | useable scale capacity | largest noted error, for this scale

scale number | material type | graduation size

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
				always zero for 1st step						

scale acceptable to | useable scale capacity | largest noted error, for this scale

scale number | material type | graduation size

step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
				always zero for 1st step						

Test Notes

PLANT CERTIFIED FOR STATE USE
INTERLOCKS O.K.
PRINTER O.K.
QUALITY SCALES UNLIMITED