



MPQP

AC, PCC, RUBBER & OTHER BATCH OPERATIONS

Report Number 11-18 calib date 06/13/2011 district 01 109 by L. Lee *WLL* plant controller Seltech AB500 sn# 2875
 plant type A/C Batch plant owner Mercer Fraser plant location Cooks Valley

scale acceptable to 350 useable scale capacity 350 largest noted error, for this scale 2

scale number	step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
1	1	100		100	always zero for 1st step	100	100	0.0	0.0	Acceptable	
	2	200		200		200	199	-1.0	-1.0	Acceptable	
	3	300		300		300	299	-1.0	-1.0	Acceptable	
	4	350		350		350	349	-1.0	-1.0	Acceptable	

scale acceptable to 4000 useable scale capacity 4000 largest noted error, for this scale 2

scale number	step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
2	1	1000		1000	always zero for 1st step	1000	1000	0	0	Acceptable	
	2	2000		2000		2000	1990	-10	-5	Acceptable	
	3	3000		3000		3000	2990	-10	-5	Acceptable	
	4	4000		4000		4000	3980	-20	-1	Acceptable	

scale acceptable to useable scale capacity largest noted error, for this scale

scale number	step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
	1				always zero for 1st step						

scale acceptable to useable scale capacity largest noted error, for this scale

scale number	step number	applied test weight	applied build-up weight	total applied weight	adjustment from last step	adjusted total weight	indicated weight	actual error	error in grads	USEABILITY	remarks
	1				always zero for 1st step						

Test Notes

PLANT CERTIFIED FOR STATE USE
 INTERLOCKS OK